

☐ SECRET☒ CONFIDENTIAL 02773☐ UNCLASSIFIEDApproved For Release 2005/05/02 : CIA-RDP78B04770A002100030002-5
CONTRACT INSPECTION REPORT

TO:

CONTRACT ADMINISTRATION & SETTLEMENT
BRANCH/PD/OL

DATE

19 October 1967

INSPECTION REPORT NO. (If final, so state)

1

ESTIMATED COMPLETION DATE

1 December 1967

NAME OF CONTRACTOR

Declass Review by NGA.

Repair of 552A-102 Fiber-Optic Viewer

THE CONTRACTOR IS ON SCHEDULE

☐ YES☒ NOTHE CONTRACTOR WILL PROBABLY REMAIN WITHIN ALLOCATED FUNDS ☒ YES ☐ NO IF ANSWER IS "NO" ADVISE RECOMMENDATION AND/OR ACTION OF SPONSORING OFFICE, ON REVERSE HEREOF. IF KNOWN, INDICATE MAGNITUDE OF ADDITIONAL FUNDS INVOLVED.

PER CENT OF WORK COMPLETED - 0%

PER CENT OF FUNDS EXPENDED - Fixed Price

HAS AN INTERIM REPORT, FINAL REPORT, PROTOTYPE, OR OTHER END ITEM BEEN RECEIVED FROM THE CONTRACTOR DURING THE PERIOD? ☐ YES ☒ NO (If yes, give details on reverse side.)HAS GOVERNMENT-OWNED PROPERTY BEEN DELIVERED TO CONTRACTOR DURING THIS PERIOD? ☐ YES ☒ NO (If yes, indicate items, quantity, and cost on reverse side.)

INCENTIVES

IS THIS AN INCENTIVE CONTRACT
IF YES, CHECK TYPE☐ YES☒ NO☐ COST☐ AWARD
FEE☐ PERFORMANCE☐ DELIVERYNOTE:
USE REVERSE SIDE FOR COMMENTS.
FINAL REPORT MUST CONTAIN INCENTIVE EVALUATION.

OVERALL PERFORMANCE OF CONTRACTOR

1. ☐ OUTSTANDING4. ☐ ABOVE AVERAGE7. ☐ UNSATISFACTORY2. ☐ EXCELLENT5. ☒ AVERAGE3. ☐ VERY GOOD6. ☐ MINIMUM ACCEPTABLE

IF OVERALL PERFORMANCE OF CONTRACTOR IS UNSATISFACTORY OR MINIMUM ACCEPTABLE INDICATE REASONS ON REVERSE SIDE.

RECOMMENDED ACTION

☒ CONTINUE AS PROGRAMMED☐ WITHHOLD PAYMENT PENDING
SATISFACTORY PERFORMANCE☐ TERMINATE☐ OTHER (Specify)

IF TERMINATION IS RECOMMENDED OR IF THIS IS A FINAL REPORT PUT COMMENTS ON REVERSE IN NARRATIVE FORM ON CONTRACTOR'S PERFORMANCE AND CERTIFY THAT ALL DELIVERABLE ITEMS UNDER THE CONTRACT HAVE BEEN RECEIVED. THESE INCLUDE, WHERE APPLICABLE, THE FOLLOWING:

ITEM	REC'D	DOES NOT APPLY	ITEM	REC'D	DOES NOT APPLY
PROTOTYPES			MANUALS		
DRAWINGS AND SPECIFICATIONS			FINAL REPORT		
PRODUCTION AND/OR OTHER END ITEMS			SPECIAL TOOLING		
			OTHER GOVERNMENT PROPERTY		

DATE OF LAST CONTACT WITH CONTRACTOR

19 October 1967

SIGNATURE OF INSPECTOR

DIVISION

☐ SECRET

☐ CONFIDENTIAL

☐ UNCLASSIFIED

Approved For Release 2005/05/02 : CIA-RDP78B04770A002100030002-5

NARRATIVE REPORT

☒ INTERIM

☐ FINAL

1. An inspection was made of the SS2A-102 Viewer and the following items were found that need repair.

a. The right objective will not transtate to the rear of the instrument from the extreme front position.

b. The braking systems on both film transports need adjustment to prevent film spillage when the transport is stopped.

c. The left 'X' counter intermittently malfunctions.

d. The center of the eyepiece field of view does not coincide with the optical axis.

e. There is electrical interference between the objective selection relays and the right 'Y' counter.

f. The left 10X - 43X objective and all of the right objectives are not centered with respect to the highest power objectives.

2. [] has been made aware of these deficiencies (verbal conversation between [] and the technical monitor on 19 October 1967) and agreed to repair the instrument during the week of 30 October 1967.

Approved For Release 2005/05/02 : CIA-RDP78B04770A002100030002-5

☐ UNCLASSIFIED

☐ CONFIDENTIAL

☐ SECRET

25X1

18 October 1967

Post Office Box 6788
Fort Davis Station
Washington, D. C. 20020

Subject: Service Contract - Fiber Optics
Stereo Viewer, Model No. 552A,
[redacted]

25X1

Reference: Your Letter Dated 1 September 1967

Dear Sir:

We are in receipt of referenced letter and have delayed answering it pending a total review of the equipment status.

As you are aware, we have been working on this equipment after normal working hours. We have done this as a courtesy to you, knowing your strict security requirements. Our review of the contract terms, our proposal, and conference notes between members of our two organizations shows that at no time was an agreement made that the technical services on [redacted] would be performed during non-working hours.

25X1

Your personnel are checking the equipment presently and, assuming only minor clean-up will be necessary, one additional trip will complete the contract. We will conduct this last visit during non-working hours provided the actions required are of a relatively simple clean-up nature. Again, this action is being taken as a courtesy to your organization, not as a contractual requirement. If we find that the additional work required is of an extensive nature, this contractor will have to request some relief, either a change in working hours or additional compensation.

It is hoped that the above clarifies our position in this matter. We understand that you are considering a service contract with [redacted] for the purpose of maintaining our equipment

25X1

CONFIDENTIAL

25X1



-2-

18 October 1967

located at your facility. It is our suggestion that if such a service contract is contemplated, we both agree beforehand on the exact working hours expected of this contractor in performing the service actions.

If we can provide additional information which you may require, please do not hesitate to contact us.

Very truly yours,



25X1

Executive Vice President

LHB/aw

CONFIDENTIAL